

Declaration of Conformity

The Directives covered by this declaration:

2014/30/EU EMC Directive
 2014/35/EU Low Voltage Directive
 2011/65/EU Restriction of the Use of certain Hazardous Substances in Electrical and Electronic Equipment (RoHS) Directive

Declares that the product(s): **Elcometer 236 Holiday Detector**

Part Number(s): **D236--15KV, D236--30KV**

Product Option(s): **T236139031, T236139032, T23613907, T23613908, T23613909, T23615550, T23622790-(1-2), T99919975, T99919988-(1-3), T99920022-(11-14), T99920022-(1-3), T99920071-(1-16), T99920082, T99920083, T99920084, T99920086, T99920252, T99920438-XX(A or B), T99920438-XXX(A or B), T99922744-XX(A or B), T99922744-XXX(A or B), T99922745-(1-10), T99922746, T99922747, T99922751, T99922752, T99922768, T99922907, T999266(21-23), T9992673(1-4), T99937(66 - 81-)**


This declaration of conformity is issued under the sole responsibility of Elcometer Limited. The products identified above comply with the requirements of the above EU Directive by meeting the following standards:

EN 61326-1:2013 IEC 61326-1:2012 Class A¹, Group 1² ISM	Electrical equipment for measurement, control and laboratory use – EMC requirements Part 1 General requirements.
EN 61010-1:2010 IEC 61010-1:2010	Safety requirements for electrical equipment for measurement, control, and laboratory use Part 1: General requirements
EN 61010-031:2002 +A1:2008 IEC 61010-031:2002 +A1:2008	Safety requirements for electrical equipment for measurement, control, and laboratory use Part 031: Safety requirements for hand-held probe assemblies for electrical measurement and test
EN 63000:2018	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

¹ Class A product: Suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.

² Group 1 ISM product: Product in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.



Signed: 
 M. C. Sellars

Manchester, UK
 Date: 16th November 2021
 Authority: Managing Director